



LEMBAGA GETAH MALAYSIA
MALAYSIAN RUBBER BOARD
RUBBER RESEARCH INSTITUTE OF MALAYSIA



MS ISO/IEC 17025

TESTING
SMM NO.069

Unit Pencirian Bahan

260 Jalan Ampang, 50450 Kuala Lumpur.
Peti Surat 10150, 50908 Kuala Lumpur, Malaysia.
Tel: (6)03-92062000, 92063681 Fax: (6)03-42519213
Email: kartinia@lgm.gov.my

Our report no. : **LGM/BTK/UPB/5.10/CP/0904/0651-0656**
Your reference no. :-

May 13, 2009

Sanctuary Health Sdn. Bhd.
464, Jalan Kuala Kangsar
30010 Ipoh
Perak DR, Malaysia

Fax : 605-2921739

Attn: Chang Chaur Long

ANALYTICAL TEST RESULTS

We report the following results:

Your letter dated 03/04/2009
Date of sample received 13/04/2009
Date of testing 08/05/2009
Test method In House Method Following ASTM D6124-01
Brand Sanctband
Lot no 051208
Measurement 0.25mm x 15cm x 1.5m
Quantity Tested 4 pcs

No.	Lab. Reference No.	Sample Reference	Residual Powder Content (mg/sample*)
3.	0653	Colour Green	0.26

* Values obtained are for whole piece of sample

“BERKHIDMAT UNTUK NEGARA”

Kartini Alias
Research Officer
Material Characterization Unit
Malaysian Rubber Board
For Director General