



LEMBAGA GETAH MALAYSIA
MALAYSIAN RUBBER BOARD
RUBBER RESEARCH INSTITUTE OF MALAYSIA

Unit Pencirian Bahan
 260 Jalan Ampang, 50450 Kuala Lumpur.
 Peti Surat 10150, 50908 Kuala Lumpur, Malaysia.
 Tel: (6)03-92062000, 92063681 Fax: (6)03-42519213
 Email: kartinia@lgm.gov.my




MS,ISO/IEC 17025
 TESTING
 SMM NO.069

Our report no. : **LGM/BTK/UPB/5.10/CP/0902/0032**
 Your reference no. :-

February 13, 2009

Sanctuary Health Sdn. Bhd.
464, Jalan Kuala Kangsar
30010 Ipoh
Perak DR, Malaysia



Fax : 605-2921739

Attn: Chang Chaur Long



ANALYTICAL TEST RESULTS

We report the following results:

Your letter dated 23/01/2009
 Date of sample received 03/02/2009
 Date of testing 12/02/2009
 Test method In House Method Following ASTM D6124-01
 Type of Sample Sanctuary Dam
 Lot no 181205
 Size 0.18mm x 5" x 5"
 Quantity Tested 4 pcs

No.	Lab. Reference No.	Sample Reference	Residual Powder Content (mg/glove)
1.	0032	Colour Green	0.10

* Values obtained are for whole piece of sample

"BERKHIDMAT UNTUK NEGARA"

Kartini Alias
Research Officer
Material Characterization Unit
Malaysian Rubber Board
For Director General

The uncertainties are for a confidence probability of no less than 95%

This report is issued in accordance with the conditions of accreditation granted by SMM which has assessed the capability of the laboratory and its traceability to recognised national standards and to the units of measurement realised at the corresponding national standards laboratory. Copyright of this report is owned by the issuing laboratory and may not be reproduced other than in full except with prior written approval of the Head of the issuing laboratory.